

OCT 26 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130493 ✓
Name	PATRECK SMITH	ID#	6397 ✓
Agency	CLEAR AFS SECURITY FORCES	Date	29 SEP 10 ✓
Instrument Location	BLDG 37, CLEAR AFS, AK 99704	Phone #	585-6093
Alco S/N	X124609 ✓	Target Value	-081 ✓
Alco Test Values	-075 ✓ 1st Alco	-075 ✓ 2nd Alco	High Pressure 1125
Signature		BMB 12/8/10	

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 16th day of DEC, 2010.

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130493

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

SEPTEMBER 29, 2010

OPERATOR'S NAME:

SMITH/PATRICK/J

OPERATOR'S NUMBER: 6397

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567

DEPT/AGENCY: CLR4

CASE/REPORT: 1234567

TEST TYPE: V

ALCO TARGET VALUE: .081

ALCO S/N: X124609

— BREATH ANALYSIS —

.081 ADJUSTED FOR 29.16 in

ALCO TARGET .078 18:12

BLANK TEST .000 18:13

INTERNAL STANDARD VERIFIED 18:13

ALCO TV 29.18 in .075 18:13

BLANK TEST .000 18:14

SUBJECT SAMPLE .000 18:14

BLANK TEST .000 18:15

ALCO TV 29.18 in .075 18:15

BLANK TEST .000 18:16

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

SEPTEMBER 29, 2010

TIME 18:09

— DIAGNOSTIC CHECK

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 012/09

HEATERS

SAMPLE CHAMBER:

BREATH TUBE:

BAROMETER: 1.18 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ^_`abcde fghijklmno
pqrstuvwxyz{|}~